



March 14, 2025

William Kotas
Intertek PSI
17 British American Boulevard
Latham, NY 12110

RE: Project: BETHLEHEM CSD HIGH SCHOOL
Pace Project No.: 70341955

Dear William Kotas:

Enclosed are the analytical results for sample(s) received by the laboratory on March 08, 2025. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in cursive script that reads "Lori Beyer".

Lori A. Beyer
lori.beyer@pacelabs.com
516-370-6014
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

Florida Certification #: E871198

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE SUMMARY

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

Lab ID	Sample ID	Matrix	Date Collected	Date Received
70341955001	HSKITCHENKETTLE2	Drinking Water	03/07/25 06:47	03/08/25 08:20
70341955002	HSA 109S	Drinking Water	03/07/25 06:41	03/08/25 08:20
70341955003	HSA 109M	Drinking Water	03/07/25 06:41	03/08/25 08:20

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE ANALYTE COUNT

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

Lab ID	Sample ID	Method	Analysts	Analytes Reported
70341955001	HSKITCHENKETTLE2	EPA 200.8	JJS	1
70341955002	HSA 109S	EPA 200.8	JJS	1
70341955003	HSA 109M	EPA 200.8	JJS	1

PACE-MV = Pace Analytical Services - Melville

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

Sample: HSKITCHENKETTLE2		Lab ID: 70341955001		Collected: 03/07/25 06:47		Received: 03/08/25 08:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	26.7	ug/L	1.0	1		03/13/25 17:33	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

Sample: HSA 109S		Lab ID: 70341955002		Collected: 03/07/25 06:41		Received: 03/08/25 08:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.0	ug/L	1.0	1		03/13/25 17:34	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BETHLEHEM CSD HIGH SCHOOL
Pace Project No.: 70341955

Sample: HSA 109M		Lab ID: 70341955003		Collected: 03/07/25 06:41		Received: 03/08/25 08:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		03/13/25 17:36	7439-92-1		

REPORT OF LABORATORY ANALYSIS



QUALITY CONTROL DATA

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

QC Batch: 389128

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70341955001, 70341955002, 70341955003

METHOD BLANK: 2045252

Matrix: Water

Associated Lab Samples: 70341955001, 70341955002, 70341955003

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	03/13/25 17:14	

LABORATORY CONTROL SAMPLE: 2045253

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	47.8	96	85-115	

MATRIX SPIKE SAMPLE: 2045255

Parameter	Units	70341942009 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	100	116	116	70-130	

MATRIX SPIKE SAMPLE: 2045257

Parameter	Units	70341942010 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	100	112	112	70-130	

SAMPLE DUPLICATE: 2045254

Parameter	Units	70341942009 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

SAMPLE DUPLICATE: 2045256

Parameter	Units	70341942010 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: BETHLEHEM CSD HIGH SCHOOL

Pace Project No.: 70341955

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70341955001	HSKITCHENKETTLE2	EPA 200.8	389128		
70341955002	HSA 109S	EPA 200.8	389128		
70341955003	HSA 109M	EPA 200.8	389128		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/Log/In Label Here



WO#: 70341955



70341955

70341955

[illegible]

WO#: 70341955

Due Date: 03/24/25

Client Name:

Project

PM: LAB

CLIENT: INTER-LATHAM

Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☐ Client ☐ Commercial ☒ Pace ☐ Other

Tracking #:

Custody Seal on Cooler/Box Present: ☐ Yes ☒ No Seals intact: ☐ Yes ☒ No Temperature Blank Present: ☐ Yes ☒ No
Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☒ Ziploc ☐ None ☐ Other Type of Ice: ☒ Wet ☐ Blue ☐ None

Thermometer Used: TH211 Correction Factor: 0.2 ☐ Samples on ice, cooling process has begun
Cooler Temperature(°C): 0.9 Cooler Temperature Corrected(°C): 1.1 Date/Time 5035A kits placed in freezer _____
Temp should be above freezing to 6.0°C

USDA Regulated Soil (☒ N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or
VA (check map)? ☐ Yes ☒ No

Did samples originate from a foreign source including Hawaii and Puerto Rico? ☐ Yes ☒ No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: ADF 3/8/25

		COMMENTS:
Chain of Custody Present:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1
Chain of Custody Filled Out:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2
Chain of Custody Relinquished:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3
Sampler Name & Signature on COC:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5
Short Hold Time Analysis (<72hr):	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	6
Rush Turn Around Time Requested:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	7
Sufficient Volume. (Triple volume provided for MS/MSD)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8
Correct Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9
-Pace Containers Used	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10
Containers Intact:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	11
Filtered volume received for Dissolved tests	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	Note if sediment is visible in the dissolved container
Sample Labels match COC	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	12
-Includes date/time/ID/Analysis Matrix: SL WT OIL OTHER		

Date and Initials of person checking preservation: ADF 3/8/25

All containers needing preservation have been	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13	<input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
pH paper Lot # <u>213624</u>			Sample #
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, NAOH > 12 Cyanide)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A		
Exceptions VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water)			
Per Method, VOA pH is checked after analysis			
Samples checked for dechlorination:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14	
Ki starch test strips Lot #			
Residual chlorine strips Lot #			
SM 4500 CN samples checked for sulf	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	15	Positive for Res. Chlorine? Y N
Lead Acetate Strips Lot #			Positive for Sulfide? Y N
Headspace in ALK Bottle (>6mm)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16	
Headspace in VOA Vials (>6mm)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	17	
Trip Blank Present	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A		
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A		

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted:

Date/Time:

Comments/ Resolution:

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.